

**Notice Information: - Advisory  
19 December 2011**

**Part 1. Product Information**

- a) Title:
- b) Product Name/Type:
- c) Reference:
- d) Manufacturer/Supplier:

**Part 2. Target Audience**

a) Target Audience:

Accident & Emergency Departments  
Chemotherapy Nurses / Consultants  
Day Surgery / Endoscopy Units  
Directors of Anaesthetics  
Family Planning Clinics  
General Practitioners  
Health Service Executive  
Hospital Chief Executive Officers  
Medical Directors  
Nursing Executive Directors  
Laboratory Managers  
Obstetrics and Gynaecology Consultants / Departments  
Oncology / Radiotherapy Departments  
Pharmacy Departments  
Point of care Managers / Coordinators  
Procurement Departments  
Relevant Wards  
Risk Managers  
Theatre Managers  
X-Ray Departments

### Part 3. Problem/Issue

a) Problem/Issue:

Lots 028619, 028924, 029997, 030240, 030431, 030602, 030826, 030998, 031332, 031460, 031662, 031798, 031924, 032106 and 032180 of the Clinitest hCG pregnancy test kits have been recalled by the manufacturer due to the potential for borderline or false positive results with commercially available controls and patient samples.

### Part 4. Background Information

a) Background Information:

The Clinitest hCG pregnancy test kit is intended for professional use only in hospitals / clinics and is not sold in pharmacies for home use. The Clinitest hCG pregnancy test may be carried out to:

- a) Determine pregnancy
- b) As part of the investigation of certain patient symptoms
- c) To rule out pregnancy before treatments or procedures that are contraindicated or should be avoided by pregnant women e.g. diagnostic x-rays.

Siemens Healthcare Diagnostics are recalling 15 lots (listed above) of Clinitest hCG pregnancy test kits following confirmation of potential borderline or false positive results with commercially available negative controls and patient samples.

## Part 5. Action to be taken

a) Action to be taken:

**ACTION OR RECOMMENDATIONS FOR HEALTHCARE PROFESSIONALS**

- Ensure that all relevant staff in your institution are informed of this recall

- Determine if you have lots 028619, 028924, 029997, 030240, 030431, 030602, 030826, 030998, 031332, 031460, 031662, 031798, 031924, 032106, and 032180 of the above product

- Locate and cease using product from the lots listed above

- Determine how much of this product has been used

- Follow the distributor / manufacturers recommendations for quarantine

and disposal of product

- Follow up with patients as required.

**ACTION OR RECOMMENDATIONS FOR PATIENTS**

- Patients who have had a recent pregnancy test administered by a healthcare

professional but are concerned that the positive result received may be incorrect should contact or seek a retest from their GP, family planning clinic or pharmacist.

## Part 6. Enquiries

- a) All enquiries should be made to:

Enquiries to the distributor should be addressed to:

Cruinn Diagnostics Ltd,  
5b/6b Hume Centre,  
Parkwest Industrial Estate,  
Nangor Road,  
Dublin 12  
Ireland

Telephone: +353-1-6297400

Fax: +353-1-6297401

E-mail: sean.mcgeown@cruinn.ie

Website: www.cruinn.ie

Enquiries to the manufacturer should be addressed to:

Siemens Healthcare Diagnostics Ltd,  
Sir William Siemens Square,  
Frimley,  
Camberley GU16 8QD,  
England

Telephone: 0044 (0) 1908 487600

Fax: 0044 (0) 1908 487601

Email: Anthony.walsh@siemens.com

All adverse incidents relating to a medical device should be reported to

All adverse incidents relating to a medical device should be reported to the:

Irish Medicines Board

Kevin O'Malley House

Earlsfort Centre

Earlsfort Terrace

Dublin 2

Telephone: +353-1-6764971

Fax: +353-1-6344033

E-mail: [vigilance@imb.ie](mailto:vigilance@imb.ie)

website: [www.imb.ie](http://www.imb.ie)

[Please click here to download a PDF version of this Safety Notice](#)