



**S
A
F
E
T
Y
N
O
T
I
C
E**

Ondal Acrobat 2000 (AC2000)

IMB Safety Notice: SN2011(07)

Circulation Date: 12 May 2011

MANUFACTURER/SUPPLIER

Various manufacturers

TARGET GROUPS

General Surgeons
Theatre and Nursing Staff
Procurement Managers
Nursing Managers
Consultant General Surgeons
Hospital Managers / CEOs
Clinical Directors
Risk Managers
Medical Device Distributors
Clinical / Biomedical Engineers

ISSUE

Ondal has become aware that there is the risk that a welded seam at the pivot of the Acrobat 2000 spring-loaded arms could break. The Acrobat 2000 spring arm has been used by a number of different manufacturers to support operating lights and monitors which have been placed on the market in Ireland.

BACKGROUND

Ondal has confirmed that the affected spring arms were manufactured from January 2000 to June 2006. The affected serial numbers may be identified by the following convention:

0111 **010** XXXXX to 0111 **066** XXXXX

010 to 066 = month **01** / year 2000 to month **06** / year 2006
XXXXX = 5-digit serial number

As outlined above, the Acrobat 2000 spring arm is used to support operating lights and monitors which have been placed on the market by various manufacturers. A number of manufacturers have initiated / completed field safety corrective actions in Ireland to address this issue. A listing of these manufacturers is provided below (please note that this is not an exhaustive list):

Trumpf Medizin Systeme – Manfred Fehn Ph: +49 3671 586 41211

Brandon Medical – Graeme Hall Ph: +44 113 277 7393

Maquet – Gerry O’Brien Ph: + 353 1 426 0032

S A F E T Y

NOTICE

Drager – Doug Sims Ph: +44 1442 213542

Gebrüder MARTIN – Sven Zehnder Ph: + 49 7461 706 453

The Irish Medicines Board (IMB) wishes to highlight these actions and requests that users remain vigilant for potentially affected devices in their facility.

ACTION OR RECOMMENDATIONS

The IMB advises that users:

- Identify the affected spring arms in your facility and contact the appropriate manufacturer for further guidance.
- Follow the manufacturer's recommendations as outlined in the associated field safety notices.
- Users are advised to exercise caution when moving or repositioning devices connected to these spring arms pending upgrade / inspection by the manufacturer.

ENQUIRIES

All adverse incidents relating to a medical device should be reported to the:

Irish Medicines Board
Kevin O'Malley House
Earlsfort Centre
Earlsfort Terrace
Dublin 2

Telephone: +353-1-6764971
Fax: +353-1-6344033
E-mail: vigilance@imb.ie
Website: www.imb.ie