


Safety Notice

Medical Devices

CRIT-LINE® Blood Chamber

Priority 2 – Warning



HPRA Safety Notice: SN2015(15)

Issue Date: 18 June 2015

MANUFACTURER / SUPPLIER	HPRA CASE REFERENCE
Fresenius Medical Care Ltd	V24356

ISSUE
The Health Products Regulatory Authority (HPRA) has been notified by Fresenius USA Manufacturing Inc Renal Therapies Group of the recall of all lots for Product Code CL10021021 of the CRIT-LINE® Blood Chamber, due to possibility of blood chamber connection leaks associated with the use of CRIT-LINE® Blood Chamber.

ACTION OR RECOMMENDATIONS
The HPRA advise that users:
1 Immediately examine your stock for CRIT-LINE® Blood Chambers.
2 If any product with product code CL10021021 is found, discontinue use immediately. A listing of all lot numbers is included in the Field Safety Notice (FSN).
3 Follow the actions outlined by Fresenius Medical Care in the FSN.
4 Forward this Safety Notice to other departments or facilities, and to all those that need

to be aware of this recall and any centres/organisation where devices may have been transferred.

- 5 Contact Fresenius Medical Care if you have any additional concerns or questions.

TARGET GROUPS

Hospital Managers / CEOs
Risk Managers
Clinical Directors
Clinical Engineers
Nursing Managers
Nursing staff
Purchasing Managers
Palliative Care Units
Intensive Care Units
Anaesthetic Officers
Nephrology Departments

Adult intensive care units
Day surgery units
Paediatric wards
Renal centres
Oncology units
Paediatric intensive care units
Neonatal units
Theatres
Healthcare professionals who use these devices

BACKGROUND

Fresenius Medical Care Renal Therapies Group, LLC received an increased number of complaints relating to blood chamber connection leaks during user of the CRIT-LINE Blood Chamber.

The manufacturer has outlined that the health consequences from blood loss may include low blood pressure, dizziness, shortness of breath, confusion, chest pain, loss of consciousness, shock, and in extreme cases, excessive blood loss may lead to death. Separation in the connection of the blood circuit that allows blood loss could provide entry for bacteria with resultant risk of bacteria, blood stream infection, and sepsis that may lead to death.

The manufacturer has issued a FSN advising of the issue – see attached.

MANUFACTURER / AUTHORISED REPRESENTATIVE/ DISTRIBUTOR CONTACT INFORMATION

Enquiries to the **manufacturer** should be addressed to:

Fresenius USA Manufacturing, Inc.
695 North 900 West
84037-4118, Kaysville
USA

Telephone: 801-549-9224
E-mail: doug.cox@fmc-na.com

Enquiries to the **authorised representative** should be addressed to:

Medical Device Safety Service (MDSS GmbH)
Schiffgraben 41
30175 Hannover

Telephone: +49 511 6262 8630
Fax: +49 511 6262 8633
E-mail: vigilance@mdss.com

Germany

HPRA CONTACT INFORMATION

All **adverse incidents** relating to a medical device should be reported to:

Health Products Regulatory Authority
Kevin O'Malley House
Earlsfort Centre
Earlsfort Terrace
Dublin 2

Telephone: +353-1-6764971
Fax: +353-1-6344033
E-mail: devicesafety@hpra.ie
Website: www.hpra.ie