

Safety Notice

Medical Devices

Klerpack BD Syringe Multi-Pack

Priority 1 – For Immediate Action



HPRA Safety Notice: SN2016(26)

Issue Date: 12 September 2016

MANUFACTURER / SUPPLIER	HPRA CASE REFERENCE
Ecolab Ltd.	V29189

ISSUE

Ecolab Ltd. has identified that the sterile packaging of the Klerpack BD Syringe Multi-Pack may contain small holes in the film of the pouch.

If holes are present, the package integrity is compromised and there is a risk that the syringe may no longer be sterile.

The issue affects a number of different product codes and lot numbers. Please refer to the Field Safety Notice (FSN) issued by Ecolab Ltd. for specific details.

We understand that these syringes have been supplied to the Irish market by multiple distributors.

ACTION OR RECOMMENDATIONS

The HPRA advise that users:

- 1 Follow the actions outlined by the manufacturer in the accompanying FSN.
- 2 Immediately examine your stock and quarantine affected devices. Refer to the FSN for a list of affected product codes and respective lot numbers.
- 3 Pass the manufacturer's FSN to all those who need to be aware of it within your organisation. Forward a copy of the FSN to any other organisations to whom affected devices have been transferred.
- 4 Forward a copy of this HPRA Safety Notice to all relevant personnel.
- 5 Forward a copy of this HPRA Safety Notice to any other persons / organisations where these devices have been transferred.
- 6 Report any adverse events / incidents associated with these devices to the manufacturer and the HPRA.

TARGET GROUPS

Accident & Emergency Departments	Laboratory directors / managers
Biomedical Engineers	Laboratory technicians
Clinical / Medical Directors	Nursing staff
Clinical Nurse Managers	Outpatient departments
Community pharmacists	Paramedics / advanced paramedics
Community users	Community pharmacies / pharmacists
Critical Care units / Intensive Care units	Private ambulance services
Day surgery units	Product compounding centres / facilities
General public	Public and private hospitals
Healthcare professionals who use these devices	Public Health nurses
Hospital managers / CEOs	Purchasing managers
Hospital personnel	Risk managers
Hospital pharmacies / pharmacists	Stores / supplies managers
HSE National Ambulance Service	Theatres
Infusion clinics / Infusion Day units	Voluntary / auxiliary ambulance services

BACKGROUND

Ecolab Ltd. identified during internal testing that small holes may be present in the film of the pouch of the device packaging.

Please refer to the manufacturer's FSN for details of specific product codes and lot numbers affected.

The HPRA is issuing this safety notice to raise awareness of this issue.

MANUFACTURER / AUTHORISED REPRESENTATIVE/ DISTRIBUTOR CONTACT INFORMATION

Enquiries to the **manufacturer** should be addressed to:

Ecolab Ltd.
PO Box 11, Winnington Avenue
Northwich, Cheshire
CW8 4DX
Great Britain

Customer Service, Ecolab Ltd
Telephone: +44 (0) 2920 854 390
Fax: +44 (0) 2920 854 391
E-mail: info@ecolab.com
Website: <http://ecolabcc.com/index.php>

HPRA CONTACT INFORMATION

All **adverse incidents** relating to a medical device should be reported to:

Health Products Regulatory Authority
Kevin O'Malley House
Earlsfort Centre
Earlsfort Terrace
Dublin 2

Telephone: +353-1-6764971
Fax: +353-1-6344033
E-mail: devicesafety@hpra.ie
Website: www.hpra.ie