

# Safety Notice

## Medical Devices

### ID-DiaCells (Multiple Products)

#### Priority 2 – Warning



HPRA Safety Notice: SN2017(10)

Issue Date: 24/02/2017

MANUFACTURER / SUPPLIER	HPRA REFERENCE
DiaMed GmbH	V30834

#### ISSUE

The HPRA has received reports from users in Ireland regarding intermittent non-specific reactions produced by ID-DiaCells. Non-specific cell reactions can result in suspect false-positive results leading to a delay in determining irregular antibody identification. This issue may also result in repeat testing.

The manufacturer issued a customer communication on 20<sup>th</sup> January 2017. A field safety notice (FSN) was subsequently issued on 10<sup>th</sup> February 2017 informing users of this issue. A follow-up FSN has also been issued by the manufacturer, providing additional information regarding this issue. Refer to accompanying FSNs for further information.

#### ACTION OR RECOMMENDATIONS

The HPRA advise that users:

- 1 Refer to the accompanying FSNs and follow the instructions provided.
- 2 Forward a copy of this Safety Notice and the accompanying FSNs to all relevant personnel.
- 3 Forward a copy of this Safety Notice and the accompanying FSNs to any other persons/organisations where these devices have been transferred

- 4 Report any adverse events / incidents associated with these devices to the manufacturer and the HPRA.

#### TARGET GROUPS

Chief Medical Scientists Clinical Directors Clinical Nurse Managers Consultant Haematologists Hospital Managers / CEOs Hospital personnel	Blood Establishments Laboratory Directors / Managers Laboratory Technicians Public and private hospitals Risk managers Stores managers
--	---

#### BACKGROUND

Unsatisfactory cell reactions have been observed in a number of different laboratories for a range of test panels and lot numbers. Unsatisfactory cell reactions may be recorded erroneously on instruments as false-positives.

The manufacturer's investigation into this issue is ongoing. The HPRA is issuing this safety notice to raise awareness of the FSNs.

The HPRA may provide an update on this issue, if further information becomes available.

#### MANUFACTURER / AUTHORISED REPRESENTATIVE / DISTRIBUTOR CONTACT INFORMATION

Enquiries to the **manufacturer** should be addressed to:

DiaMed GmbH Pra rond 23 1785 Cressier FR Switzerland	Telephone: +41-26-674 5236 Fax: +41-26-674 5469 E-mail: <a href="mailto:RA-request_Cressier@bio-rad.com">RA-request_Cressier@bio-rad.com</a> Website: <a href="http://www.diamed.com">www.diamed.com</a>
---	---

Enquiries to the **distributor** should be addressed to:

Fannin Fannin House South County Business Park Leopardstown, Dublin 18.	Telephone: +353-1-290 7000 E-mail: <a href="mailto:quality@fannin.eu">quality@fannin.eu</a> Website: <a href="http://www.fannin.eu">www.fannin.eu</a>
---	---

#### HPRA CONTACT INFORMATION

All **adverse incidents** relating to a medical device should be reported to:

Health Products Regulatory Authority  
Kevin O'Malley House  
Earlsfort Centre  
Earlsfort Terrace  
Dublin 2

Telephone: +353-1-676 4971  
Fax: +353-1-634 4033  
E-mail: [devicesafety@hpra.ie](mailto:devicesafety@hpra.ie)  
Website: [www.hpra.ie](http://www.hpra.ie)