

## Urgent Field Safety Notice

**Product Names: BATH CHAIR AND BATH CORNER CHAIR**

**FSCA Identifier: 2017/003/016/401/001**

**Type of Action: Preventative**

Date:	05 <sup>th</sup> April 2017
Attention:	All end users of Smirthwaite Ltd. Bath Chairs and Bath Corner Chairs.
Details of affected devices:	Bath Chairs part numbers 7607 and 7609; Bath Corner Chairs part numbers 7511 and 7512 with the following serial numbers 126560, 132862 to 132995, 133009 to 133995, 134011 to 134993 and 135001 to 135516.
Description of problem:	The pommel may become detached from its carrier exposing a rough/sharp surface which may cause injury to the user.
Advice on action to be taken by the user:	Remove the pommel and return it to Smirthwaite Ltd. Who will replace it with a new pommel.
Transmission of the Field Safety Notice:	<p>This notice and a letter to all customers of the above products.</p> <p>Please pass this Field Safety Notice to all your staff that come into contact with this product or need to be notified of this notice.</p> <p>We should be grateful if you would confirm receipt of this notice by returning the prepaid envelope or emailing "FSN Bath Chair or FSN Bath Corner Chair" to <a href="mailto:info@smirthwaite.co.uk">info@smirthwaite.co.uk</a></p>
Contact reference person:	<p>Richard Snape Managing Director Smirthwaite Ltd. 17 Wentworth Road Heathfield Newton Abbot Devon, England. TQ12 6TL 01626 835552 MHRA has been informed</p> <p>Richard Snape Managing Director Smirthwaite Ltd.</p>