

FSN Ref: 2020004

FSCA Ref: N/A

Date: 05/10/2020

Urgent Field Safety Notice
Device Commercial Name

For Attention of*: Clinical Laboratory managers and lab technicians

Contact details of local representative (name, e-mail, telephone, address etc.)*

This could be a distributor or local branch of the manufacturer. To be added at the appropriate stage in the different local languages
--

Urgent Field Safety Notice (FSN)

0318P KWIK-STIK™ 2 Pack Clostridium perfringens derived from ATCC® 13124™

5190P QC Sets and Panel ANC Comprehensive QC Set

Risk addressed by FSN

1. Information on Affected Devices*	
1.	1. Device Type(s)* Unassayed quality control material for microbiology assays.
1.	2. Commercial name(s) 0318P KWIK-STIK™ 2 Pack Clostridium perfringens derived from ATCC® 13124™ 5190P QC Sets and Panel ANC Comprehensive QC Set
1.	3. Unique Device Identifier(s) (UDI-DI) 0318P UDI: 20845357006213 5190P UDI: 70845357030626
1.	4. Primary clinical purpose of device(s)* KWIK-STIK™ and LYFO-DISK™ microorganisms are intended to be used as controls to verify the performance of assays, reagents or media that are intended to be used in microbial testing for the detection and identification of a cultured microorganism isolate. Each KWIK-STIK contains a qualitative lyophilized microorganism pellet, ampoule of hydrating fluid and inoculating swab. Everything you need to grow reference cultures for QC testing is included in this one handy device. Each LYFO-DISK™ contains 6 lyophilized pellets for flexibility in the lab. The products are unassayed, meaning it is not intended to be used with any specific assay. 0318P KWIK-STIK™ 2 Pack Clostridium perfringens derived from ATCC® 13124™ 5190P QC Sets and Panel ANC Comprehensive QC Set contains two KWIK-STIKs of each strain listed below (14 KWIK-STIKs total). This set contains 0318P as one component: 0585P Bacteroides ovatus derived from ATCC® BAA-1296™ 0445P Bacteroides vulgatus derived from ATCC® 8482™ 0318P Clostridium perfringens derived from ATCC® 13124™ 0586P Clostridium septicum derived from ATCC® 12464™ 0331P Paenoclostridium sordellii derived from ATCC® 9714™ 0583P Corynebacterium striatum derived from ATCC® BAA-1293™ 0584P Parabacteroides distasonis derived from ATCC® BAA-1295™
1.	5. Device Model/Catalogue/part number(s)* 0318P, 5190P
1.	6. Software version N/A
1.	7. Affected serial or lot number range

FSN Ref: 2020004

FSCA Ref: N/A

	0318P Lot: 318-234-4
	5190P Lot: 5190-08
1.	8. Associated devices
	N/A

2. Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem*
	Mislabeled foil pouch that the individual KWIK-STIK™ is placed in; the foil pouch is labeled as lot 218-234-4. All other labels (KWIK-STIK™ device, KWIK-STIK™ canister, and Certificate of Analysis are correct.
2.	2. Hazard giving rise to the FSCA*
	All foil pouches for lot 318-234-4 have the lot number misprinted as 218-234-4. The canister label is correct, and the label on the actual KWIK-STIK is correct. The KWIK-STIK label is detached by the user and used to identify the plate. If the end-user stores the KWIK-STIKs in the canister and opens the pouch without noticing the pouch label is incorrect, the user would not be affected.
2.	3. Probability of problem arising
	All customers that have received this lot may experience this hazard.
2.	4. Predicted risk to patient/users
	These products are controls for diagnostic assays (but they are not diagnostics themselves). There is no health risk posed by this non-conformance. Since the KWIK-STIK™ device label and KWIK-STIK™ canister labels are correct, the end-user may not noticed the mislabeled foil pouch. If the end-user stores the KWIK-STIK™ outside of the canister, they may notice the misprinted lot number on the foil pouch, but the KWIK-STIK™ device label will be correct on the inside. The KWIK-STIK™ device label is what customers tear off to label their agar plate with.
2.	5. Further information to help characterize the problem
	N/A
2.	6. Background on Issue
	N/A
2.	7. Other information relevant to FSCA
	N/A


FSN Ref: 2020004

FSCA Ref: N/A

	3. Type of Action to mitigate the risk*	
3.	1. Action To Be Taken by the User* <input checked="" type="checkbox"/> Identify Device <input type="checkbox"/> Quarantine Device <input type="checkbox"/> Return Device <input type="checkbox"/> Destroy Device <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Follow patient management recommendations <input type="checkbox"/> Take note of amendment/reinforcement of Instructions for Use (IFU) <input checked="" type="checkbox"/> Other <input type="checkbox"/> None Use or discard the affected products depending on your lab procedures and how this information affects your usage of the product.	
3.	2. By when should the action be completed?	Upon receipt of this notice
3.	3. Particular considerations for: N/A Is follow-up of patients or review of patients' previous results recommended? No	
3.	4. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes
3.	5. Action Being Taken by the Manufacturer <input type="checkbox"/> Product Removal <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input checked="" type="checkbox"/> Other <input type="checkbox"/> None Quarantine all current stock and initiate FSCA	
3	6. By when should the action be completed?	Completed
3.	7. Is the FSN required to be communicated to the patient /lay user?	No
3.	8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet? N/A	

FSN Ref: 2020004

FSCA Ref: N/A

4. General Information*		
4.	1. FSN Type*	New
4.	2. For updated FSN, reference number and date of previous FSN	N/A
4.	3. For Updated FSN, key new information as follows:	N/A
4.	4. Further advice or information already expected in follow-up FSN? *	No
4	5. If follow-up FSN expected, what is the further advice expected to relate to:	N/A
4	6. Anticipated timescale for follow-up FSN	N/A
4.	7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
	a. Company Name	Microbiologics, Inc.
	b. Address	200 Cooper Ave North, St. Cloud, MN 56303 USA
	c. Website address	www.microbiologics.com
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. * After a risk assessment, national competent authorities have not been notified about this communication because there is no risk of harm to patients or users.	
4.	9. List of attachments/appendices:	Customer Reply Form
4.	10. Name/Signature	Kali Sorum, Technical Support Manager
		

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organization or to any organization where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organizations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.*</p>

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional.