

URGENT MEDICAL DEVICE CORRECTION

To users of the ACUSON Juniper and ACUSON Juniper Select Ultrasound systems:

Dear Valued Customer:

This letter is to notify users of the DICOM Structured Reporting (SR) feature of a potential safety concern when exporting examination data with the Cardiac DICOM SR exam feature provided on the ACUSON Juniper and ACUSON Juniper Select Ultrasound Systems.

What is the issue?

This issue is specific to and limited to the use of the Cardiac DICOM SR feature:

- When a user uses and configures the Cardiac DICOM SR feature to display either the MINIMUM (Min) or the MAXIMUM (Max) measured value, AND
- Makes multiple measurements of the cardiac region during an examination, AND
- Exports the results into the Cardiac DICOM SR feature, THEN
- The Cardiac DICOM SR viewer will NOT display the MINIMUM (Min) or MAXIMUM (Max) value and will instead display the LAST measured value. In that case, the displayed measurement in the Cardiac DICOM SR viewer may NOT be the Minimum (Min) or MAXIMUM (Max) value measured during the examination.

What is NOT affected by this issue?

This issue does NOT affect any value displayed on the ACUSON Juniper Ultrasound System Final Report.

If you do NOT use the Cardiac DICOM SR feature, this issue does NOT affect your results.

This issue does NOT affect measurements configured to display either the AVERAGE or LAST measured value in the Cardiac DICOM SR feature.

This issue does NOT affect calculations derived from cardiac exam measurements.

This issue does NOT affect the ACUSON Juniper Ultrasound System, product version 2.5, software version VB30D.

What is the potential risk to patient safety?

Unexpected reported measurement values viewed in the Cardiac DICOM SR viewer could contribute to misdiagnosis of a patient's condition or influence patient management decisions in a negative way.

The potential risk of misdiagnosing a severe valvular or congenital heart condition from over or under represented measurements is unlikely as the cardiac exam report calculation values, the recorded ultrasound imagery, and screen-captured images are all accurately maintained when exporting to a Cardiac DICOM SR viewer.

As of February 7, 2024, Siemens Healthineers has not received any report of injuries related to this issue.

What steps can the user take to avoid the potential risk of this issue?

To avoid potential misdiagnosis by an over- or underrepresented measured value, it is recommended to refer to the measurement values displayed on the ACUSON Juniper Ultrasound System Final Report when the configuration is to display the Min or Max value.

What if I transferred and reviewed cardiac examination data from a DICOM SR tool within a clinical context?

Siemens Healthineers recommends a review of previous cardiac ultrasound examination results where the clinical assessment was reviewed using a Cardiac DICOM SR viewer.

Should an adverse reaction or quality problem be experienced with the use of this product, please report the incident to Siemens Healthineers.

How will the issue be resolved?

Siemens Healthineers will correct this issue with a free-of-charge software update to your ACUSON Juniper and ACUSON Juniper Select system.

Your Customer Service Engineer from Siemens Healthineers will contact you to schedule a facility visit to update the system or inform you of a remote update when the software update is available. The software update is currently under development and estimated to be available by spring/summer of 2024.

Dissemination of the content of this notice:

Please ensure that all users of ACUSON Juniper and ACUSON Juniper Select systems within your organization, and others who may need to be informed, receive the safety relevant information provided with this notice and take the actions specified herein.

For users in the United States of America:

If an adverse event or quality problem is experienced with the use of this product, the issue may be reported to the FDA's MedWatch Adverse Event Reporting program either online, or by regular mail or fax.

Patient safety and customer satisfaction are our highest priorities. We appreciate your cooperation with this product advisory and apologize for any inconvenience this causes your institution. If you have further questions, please contact the Siemens Healthineers Ultrasound Service Customer Care Center at 1-800-888-7436.

Sincerely,



*Electronically signed by: Jim Dabbs
Reason: I am approving this document
Date: Feb 8, 2024 13:16 PST*

James R. Dabbs
Vice President, Quality & Regulatory
Siemens Medical Solutions USA, Inc.
Ultrasound Business Area