

Urgent Field Safety Notice

ReoPro Filter kit

12 Feb 2016

Type of action - modification

Date: 12 Feb 2016

Attention: Catheter lab personnel

Details on affected devices:

Distributor	Batch Number	Expiration Date On the pack	Correct Expiration Date
<i>Eli Lilly</i>	<i>201501</i>	<i>June 2019</i>	<i>August 2017</i>

Description of the problem:

Eli Lilly and Company would like to draw your attention to the following important information regarding incorrect expiry dating on ReoPro Filter Kits that have been supplied to you with your ReoPro delivery. We are hereby notifying you that the correct expiry date for this lot is **August 2017**. There is no risk to patients at this point in time as all components are within their individual expiry dates.

Advise on action to be taken by the user:

- Check your inventory for the lot number identified.
- If this specific lot is found, please do not use beyond August 2017.
- Always ensure you check the expiry date of each component before storing and using it.

Transmission of this Field Safety Notice:

This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (If appropriate)

Please maintain awareness of this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action. (if appropriate).

Contact reference person:

Úna Walsh
Quality Leader UK and ROI
Eli Lilly and Company Ltd
Lilly House, Priestly Road,
Basingstoke, Hants
RG26 9NL
Tel – 00 44 1256 315000
Fax - 00 44 1256 356856

The undersign confirms that this notice has been notified the appropriate Regulatory Agency

Signature