

URGENT FIELD SAFETY NOTICE

ARTIS/EVOSYS DIALYSIS SYSTEM EQUIPPED WITH SW 8.09.12, 8.09.13, 8.15.06 AND 8.33.02

CORRECT HANDLING OF THE ALARM #642 “ARTERIAL CHAMBER: LEVEL ADJUSTMENT REQUIRED “

FA-2014-011

22 December 2014

Dear Customer,

We want to remind users of the importance of following the instructions in the Operator's Manual when Alarm #642 “Arterial Chamber: Level Adjustment Required” occurs on the Artis/Evosys dialysis system. Gambro has become aware of events where the instructions have not been followed and this may deactivate the alarm mechanism for the remaining treatment time.

AFFECTED DEVICES:

Artis/Evosys Dialysis System equipped with SW 8.09.12, 8.09.13, 8.15.06 and 8.33.02.

DESCRIPTION OF THE PROBLEM:

When the Alarm #642 “Arterial Chamber: Level Adjustment Required” is generated by the Artis/Evosys dialysis monitor, this is an indication that the blood level inside the arterial chamber is too low and may result in the blood pump pumping air to the dialyzer. The air bubble detection system is designed to prevent air bubbles greater than 20 µl from reaching the patient but not microbubbles below that detection threshold.

As explained in the Operator's Manual, there are a number of potential causes for the lowering of the blood level and air entering the arterial chamber, including:

- Arterial infusion lines not correctly clamped;
- Empty saline/medication bags or bottles connected to the arterial chamber;
- Saline/medication bag improperly connected to the arterial infusion lines;
- Incorrect operations or syringe defect while accessing the arterial injection port;
- Vascular access not properly connected to arterial patient line.

When Alarm #642 occurs, the Operator's Manual provides clear instructions on how to increase the blood level in the Arterial Chamber. The Operator's Manual also advises that if the RESET button is pressed without performing the blood level adjustment procedure, the Alarm will no longer be triggered during the remaining treatment time.

URGENT FIELD SAFETY NOTICE

ARTIS/EVOSYS DIALYSIS SYSTEM EQUIPPED WITH SW 8.09.12, 8.09.13, 8.15.06 AND 8.33.02

CORRECT HANDLING OF THE ALARM #642 “ARTERIAL CHAMBER: LEVEL ADJUSTMENT REQUIRED “

As indicated above, performing the treatment with blood level too low in the Arterial Chamber may result in microbubbles smaller than 20 µl reaching the patient. Although we have received no reports of long term health consequences in relation to this issue, it is important that Alarm #642 is appropriately addressed to prevent this possibility.

ACTIONS TO BE TAKEN BY THE USER:

- The actions required for resolving Alarm #642 must be executed correctly following the instructions in the Operator’s Manual.
- After the RESET button is pressed, the blood level adjustment procedure shall be performed, otherwise the alarm is deactivated for the remaining treatment time.
- Please return the attached **Customer Reply Form** to Gambro to confirm that you have received this Field Safety Notice and that you and all involved operators have been informed of this issue.

TRANSMISSION OF THIS FIELD SAFETY NOTICE:

This Notice needs to be passed on all those who need to be aware within your organisation. Please maintain awareness on this Notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.

CONTACT REFERENCE:

Please contact your local Gambro sales office/distributor. We are available to answer any questions you may have regarding this issue and the suggested actions.

Gambro is committed to maintain the continuity of care for patients relying on receiving these lifesaving treatments and to ensure the safety and effectiveness of its products.

Gambro plans to release a SW modification to prevent the possibility of continuing the treatment without following the instructions for addressing Alarm #642. We expect the modification will be available during Q2 2015.

We have advised the relevant Competent Authorities of this Field Safety Notice.

Dott.sa Raffaella Turci



Senior Manager QA/RA Monitors Operations Medolla

Gambro Dasco S.p.A
Via Modenese, 66
41036 Medolla (MO)
Italy