



«Nom\_de\_Etablissement\_»  
«Nom\_du\_correspondant»

«Adresse\_1»  
«Adresse\_2»  
«CP\_VILLE\_»

Quality contact : ☎ 00 33 139 926 351  
✉ [nusur@vygon.com](mailto:nusur@vygon.com)

Our ref : 15 1001 «N\_de\_Chrono»CO/NU

Ecouen, 30<sup>th</sup> June, 2015

**Subject :** **Field Safety Notice - Lot recall**  
**20 ml Vygon Nutrisafe C-GON Enteral syringes**  
**code 1015.202 batch G86115-2**

Dear Madam, Sir,

We have identified a non-systematic mix up of Enteral syringes in the same box for specific **lot G86115-2**. In the same box, we have found 20 ml Vygon nutrisafe C-GON enteral syringes and 20ml Codan Medical Aps Enteral CE0543 syringes. These syringes cannot be connected to our Nutrisafe devices.



**Therefore, we ask you to withdraw them.**

We ask you to kindly acknowledge receipt of this letter, to complete and return the attached form by indicating involved quantities withdrawn from your institution.

*The French health authorities (ANSM) have been informed of this recall.*

*For further information, you can directly contact M.xxxxx quality contact, or sales administration phone number xxxxxxxxxxxx.*

We apologize for any inconvenience this FSCA may cause,

**Christine OBER - Pharmacist**  
**Postmarket Quality Director**

Service Qualité  
Tél : 01.39.92.63.51  
Fax : 01.39.92.64.82  
E-mail : [quality@vygon.com](mailto:quality@vygon.com)

Service Clients France  
Tél : 01.39.92.63.81  
Fax : 01.39.90.99.37  
E-mail : [commandes@vygon.com](mailto:commandes@vygon.com)

Service Clients Export  
Tél : 01.39.92.64.17  
Fax : 01.34.29.19.34  
E-mail : [export@vygon.com](mailto:export@vygon.com)





**RECALL 20ml Vygon Nutrisafe C-GON Enteral syringes**

**CODE 1015.202 batch G86115-2**

**ACKNOWLEDGMENT AND CUSTOMER RESPONSE FORM**

**Please complete and return this form:**

**by fax: xxxxxxxxxxxx**

**or e-mail to : xxxxxxxxxxxx**

**Name and address of the institution :**

**Full name of the person to contact:** 15«N\_de\_Chrono»CO/NU «Email\_correspondant»

«Town\_»

**Function :**

**Phone number :**

**Please tick the appropriate box:**

**You have the 20ml Vygon Nutrisafe C-GON Enteral syringes code 1015.202:**

**Lot G86115-2**     Yes     NO    If you have units, number of units removed: \_\_\_\_\_

**Signature and Date :**

Service Qualité  
Tél : 01.39.92.63.51  
Fax : 01.39.92.64.82  
E-mail : quality@vygon.com

Service Clients France  
Tél : 01.39.92.63.81  
Fax : 01.39.90.99.37  
E-mail : commandes@vygon.com

Service Clients Export  
Tél : 01.39.92.64.17  
Fax : 01.34.29.19.34  
E-mail : export@vygon.com

