

To the attention of Medical Device Vigilance
responsible / Central Pharmacy

Saint Priest, October 19th, 2023

Subject: URGENT - FIELD SAFETY NOTICE – INTEGRA – Integra® Cranial Access Kit, without drugs, Bit and Guard – Reference: INS5HND – RECALL

Legal manufacturer: INTEGRA PAIN MANAGEMENT - 3498 WEST 2400 SOUTH #1050 WEST VALLEY CITY, UT 84119 – US-MF-000018493

EC Representative:

INTEGRA LIFESCIENCES (France) SAS – Immeuble Séquoia 2 – 97 Allée Alexandre Borodine – 69800 SAINT PRIEST, France – SRN : FR-AR-000002474

Medical device:

The Cranial Access Kit consists of various accessories used during a ventriculostomy procedure. The kit includes a hand drill with chuck, and a drill bit with depth guard. The depth guard allows accurate, secure, and easy to use adjustment of hand drill depth. Hand drill depth is selected by setting the depth guard to the required distance. The kit contains various instruments, including scalpels, needles, syringes, skin marker and ruler, fenestrated drape, towels, sponges and gauze.

The Cranial Access Kit is single use and disposable.

Primary clinical purpose of device:

The Cranial Access Kit allows for access to the subarachnoid space or the lateral ventricles of the brain. The kit is intended to be used with an external drainage and monitoring system in selected patients to reduce intracranial pressure (ICP), to monitor CSF, to provide temporary drainage of CSF, and to monitor ICP.

Concerned reference and lot numbers:

INS5HND - Integra® Cranial Access Kit, without drugs, Bit and Guard

Lots:

6710219

6896403

Dear Valued Integra Customer,

Integra LifeSciences is voluntarily issuing this Field Safety Notice for the recall of Integra® Cranial Access Kit part number INS5HND distributed from December 22, 2022, to September 1, 2023: see details in Table 1 below.

Product Name Unique Device Identifier (UDI)	Product Code	Lot number	Manufacturing Dates	Expiry Dates	Distribution Dates
Codman® Cranial Access Kit, without drugs, bit and guard UDI: 10381780263906	INS5HND	6710219	19/12/222	30/05/2025	From 22/12/2022 to 01/09/2023
		6896403	13/06/2023	30/10/2025	

Table 1: Product and Distribution Information

Through an internal investigation, it was identified that there is a defect in the outer packaging that houses the Cranial Access Kits impacting the two lots mentioned above. This defect can cause the packaging to split without any additional forces outside of regular manufacturing and sterilization/shipping forces being applied. The outer packaging represents the sterile barrier for the Cranial Access Kits, which are sold as single use, sterile kits.

This voluntary recall is limited to reference INS5HND and specific lots outlined in Table 1. No other products sold in EMEA are impacted. All other lots of this product may be used with confidence and without limitation.

Risks to Health

Based on the health hazard evaluation conducted for this issue, there are potential harms as infection, fever, allergic reaction, and/or toxic reaction if a kit with compromised sterility is used. The overall health hazard index for these harms is low. In addition, there are no long-range health consequences expected due to this issue. Furthermore, if the kits were used and standard post-operative care is followed, no further patient follow-up is required.

To date, no complaints have been received and no serious injuries have occurred due to this issue.

Actions to be Taken by Customers

1. Please **review and understand** the information provided in this letter.
2. If **you do have** affected kits:
 - a. Quarantine the kits immediately.
 - b. Check the box on the enclosed form “I do have affected kits.”
 - c. Record on the form the total quantity of affected kits and lot number that you have.
3. If **you do not have** affected kits, check the box, “I do not have affected kit.”
4. Please return the completed reply form by email to emea-fsca-neuro@integralife.com, or Fax to +33 (0)4.37.47. 59.30. By filling this form, you confirm that you have received this Safety Notice and you

intend to fully comply with this notification. **We expect a response within 3 weeks.** You also confirm that this notification has been forwarded to every concerned person in your organization.

5. At receipt of your form, and if it is noted that you have affected kits, Integra Customer Service will contact you and provide a Return Material Authorization (RMA) number and directions to return the affected product. A credit or replacement order can be processed based on the availability for the quantity noted on the form.
6. We recommend that you retain a copy of the form for your records.

The receipt of this form ensures that Integra has achieved a level of effectiveness in communicating this information.

National Competent Authorities may perform audits of field actions of this nature to verify that our customers have been notified and understand the nature of the field action being taken.

The National Competent Authority of your country has been alerted of this Field Safety Corrective Action.

Thank you for your cooperation with this Field Safety Corrective Action and for returning the attached Reply Form.

Please feel free to contact our Post Market Surveillance Department at emea-fsca-neuro@integralife.com for any additional questions. Your cooperation is appreciated, and we thank you for your continued collaboration.

Yours Sincerely,


Angélique AUBERT
Materiovigilance Correspondent

Appendix: Field Safety Notice Customer Reply Form (2 pages)

CUSTOMER REPLY FORM

1. Field Safety Notice (FSN) information	
FSN Reference number	FSN-2023-HHE-009
FSN Date	19/10/2023
Device name	Integra® Cranial Access Kit without Drugs
Product Code	INS5HND
Lots	6710219 / 6896403

2. Customer Details	
Account Number	
Healthcare Organisation Name*	
Organisation Address*	
Department/Unit	
Shipping address if different to above	
Contact Name*	
Title or Function	
Telephone number*	
Email*	

3. Customer action undertaken on behalf of Healthcare Organisation		
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content.*	
<input type="checkbox"/>	I performed all actions requested by the FSN *	
<input type="checkbox"/>	The information and required actions have been brought to the attention of all relevant users and executed. *	
<input type="checkbox"/>	I have checked my inventory*	
<input type="checkbox"/>	I <u>do have</u> affected kits and I have quarantined them.*	Quantity: Lot:
<input type="checkbox"/>	I <u>do not</u> have any affected units	Quantity: Lot:
<input type="checkbox"/>	I have a query please contact me	<i>Customer to enter contact details if different from above and brief description of query</i>
Print Name*		<i>Customer print name here</i>
Signature*		<i>Customer sign here</i>
Date*		

4. Return acknowledgement to Sender	
Email	emea-fsca-neuro@integralife.com
Distributor Helpline	+33 (0) 6 30 20 69 66
Postal Address	Post Market Surveillance Department Integra Immeuble Séquoia 2, 97 allée Alexandre Borodine Parc technologique de la Porte des Alpes 69800 Saint Priest, France
Web Portal	https://integralife.eu/
Fax	+33 (0)4 37 47 59 30
Deadline for returning the customer reply form*	9/11/2023

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective action.