

IMPORTANT NOTICE

Catalogue No. CQ5051, CQ5052, CQ5053

Troponin T target values are no longer quoted for the quality control materials listed above.

Troponin T values have been removed from the IFU. Randox no longer make claims for expected value or stability of the analyte in these quality control materials.

These products are no longer suitable for control of the Troponin T assay.

Authorised by: M Kelly
RA Specialist
Date: 26 Sept 2018.

Ref: 98/23457