



Field Safety Notice

NRS/Moving Solutions Slings – All types

FSN/0002 May 2011

Device Modification – Revision to Instructions for Use

Date: 10 June 2011

Attention: All customers who have purchased or use NRS and/or Moving Solutions Slings – all types.

Description of the problem:

Due to an incident involving an old Moving Solutions Sling, Nottingham Rehab Supplies are updating all of their Sling User Instructions to include improved guidelines on intended use, laundering and inspection method. The requirement to inspect the sling before each use is added together with a more detailed sling inspection procedure. Failure to read and implement the revised User Instructions could result in User Injury.

Actions to be taken by the User

All users should read, understand and communicate this Field Safety Notice and revised User Instructions to all persons involved in the use of NRS and/or Moving Solutions Slings. All slings should be inspected in accordance with the revised User Instructions. The revised User Instructions should be retained with the sling for future reference.

All customers should complete the attached Field Safety Notice Receipt Form and return this to the contact address below or fax to 0845 121 8112 within 14 days.

This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.

The undersigned confirms that this notice has been notified to the appropriate Regulatory Agency

We apologize for any inconvenience caused and thank you for your assistance in this matter.

John Turner,
Nottingham Rehab Supplies, Clinitron House, Excelsior Road. Ashby de la Zouch,
Leicestershire, LE65 1JG
Tel: 01530 418363 e-mail: jturner@nrs-uk.co.uk



Field Safety Notice Receipt Form

24 May 2011

FSN002 All Types – NRS/Moving Solutions Slings

This form is to confirm receipt of the enclosed Nottingham Rehab Supplies Field Safety Notice FSN002 dated 24 May 2011 regarding all types of NRS/Moving solutions slings and must be completed and returned within 14 days.

I confirm receipt of the above Field safety Notice. The information forwarding or action has been completed.

Date: _____

Signed : _____

Name: _____

Organisation: _____

Address: _____

Phone: _____

E-mail: _____

Please return the completed form to:

John Turner
Nottingham Rehab Supplies
Clinitron House
Excelsior Road
Ashby de la Zouch
Leicestershire
LE65 1JG

Or fax to 0845 121 8112