

**Notice Information: - Advisory
05 October 2011**

Part 1. Product Information

- a) Title:
- b) Product Name/Type:
- c) Reference:
- d) Manufacturer/Supplier:

Part 2. Target Audience

- a) Target Audience:

Part 3. Problem/Issue

- a) Problem/Issue:

Part 4. Background Information

a) Background Information:

Hill-Rom has received a number of reports of actuator failures. As a result, Hill-Rom initiated a field safety corrective action, fitting all affected lifts in use on the market with an 'Outer Tube' accessory, to prevent the actuator from collapsing in the event that it should malfunction.

There are approximately 131 mobile patient lifts on the Irish market, which have not received the 'Outer Tube' accessory. The affected products and serial numbers (sequential) for this action are, as follows:

Uno 100 EM/EE: s/n 7090001 - 7096199

Viking S: s/n 7300301 - 7301299

Uno 101: s/n 10001 - 11000

Viking XS: s/n 7400301 - 7400749

Uno 102: s/n 20001 - 21300

LikoLight: s/n 2500001 - 2505899

Uno 102 EM/EE/ES: s/n 30001 - 48099

Part 5. Action to be taken

a) Action to be taken:

Hill-Rom has been unsuccessful in their attempts to date to locate all devices in Ireland affected by this field safety corrective action.

The IMB advises that users:

1. Ensure the appropriate personnel are made aware of this notice.
2. Identify the location of all affected mobile patient lifts.
3. Determine if you / your institution have medical devices affected by this issue.
4. If you / your institution have affected medical devices, follow the manufacturer's recommendations as outlined in the attached field safety notice and ensure that corrective action is completed on all affected medical devices.

Part 6. Enquiries

- a) All enquiries should be made to:

Enquiries to the manufacturer should be addressed to:

UK Customer Care

Hill-Rom UK

Clinitron House

Ashby Park

Ashby da la Zouch

Leicestershire LE65 1JG

United Kingdom

Telephone: +44-1530-562167

Fax: +44 -1530-411555

Email: uk.customer.care@hill-rom.com

All adverse incidents relating to a medical device should be reported to the:

Irish Medicines Board

Kevin O'Malley House

Earlsfort Centre

Earlsfort Terrace

Dublin 2

Telephone: +353-1-6764971

Fax: +353-1-6344033

E-mail: vigilance@imb.ie Website: www.imb.ie

[Please click here to download](#)

Website: www.imb.ie

[Please click here to download a PDF version of Safety Notice](#)

[Please click here to download copy of Field Safety Notice](#)