



11750 12th Avenue South, Burnsville, MN 55337-1295 Phone (952) 890-1940 FAX (952) 890-0418 [www.apothecaryproducts.com](http://www.apothecaryproducts.com)

**Customer Communication: Product Recall**

**Affected Products:**

**Code AP64783 – EZ Complete Baby Care Kit**  
**Code Ap67111 - Duffy Dog Medicine Spoon**  
**Code Ap67110- Crocodile Medicine Spoon**

**Code AP67753 - Pacifier Nursing Syringe**  
**Code AP96231 - Pediatric Urine Collector**

**June 17, 2011**

Dear Pharmacist,

We write to you with reference to a recent safety notice SN2011 (15) issued by the Irish Medicines Board (IMB) in relation to the above mentioned products from Apothecary Products Inc. These products do not have a CE mark and a four digit Notified Body number which confirms the measuring aspect of the medical device has been certified by a Notified Body as required by medical devices legislation. The affected products are currently marketed without restrictions in the United States, and are listed with the US Food and Drug Administration (FDA) as medical devices. There have been no incidents reported to the manufacturer, Apothecary Products Inc. or the IMB in relation to these devices however as a precautionary measure, Apothecary Products Inc. is asking consumers to stop using these products.

What to do:

Apothecary Products Inc. has distributed some of these products through Fleming Medical, Dock Road, Limerick to pharmacies in Ireland. Fleming Medicals records show that you have purchased some or all of these products. We ask you to check your stock and see if you have still these items - if so please cease sales and contact Fleming Medical by email to [info@flemingmedical.ie](mailto:info@flemingmedical.ie) or free phone 1800 307777 – detailing the quantity you have of each item and they will arrange collection of stock and also issue you with a credit for the value of goods.

Thank you for your assistance.

Yours Sincerely

  
Jennifer Fulford  
Apothecary Products

