


Safety Notice

Medical Devices

Eleganza de Luxe

Priority 2 – Warning



HPRA Safety Notice: SN2019(05)

Issue Date: 1st February 2019

MANUFACTURER / SUPPLIER	HPRA CASE REFERENCE
Linet spol. S.r.o.	V36009

ISSUE

Certain models of Eleganza de Luxe hospital beds are equipped with footswitches for height adjustment of the mattress area. The footswitch is lockable by the operator to prevent unintended use of the footswitch. Linet, the manufacturer, has become aware of users not using the footswitch locking function, which can lead to unintended activation of the bed and potential injury to the operator or patient.

Due to the passage of time, changes in the service providers and distribution network of the beds, Linet does not have full visibility of the locations and/or users of the beds in Ireland. The HPRA is issuing this Safety Notice to raise awareness of the field safety notice (FSN).

ACTION OR RECOMMENDATIONS

The HPRA advises that healthcare professionals:

1. Refer to the accompanying FSN and follow the recommendations provided.
2. Acknowledge receipt of the FSN to the supplier / service provider for this bed.

3. Forward a copy of this safety notice and the FSN to all relevant personnel within your organisation and to any other organisations/persons to which/whom these devices have been transferred.
4. Report any concerns regarding these devices and this issue to the manufacturer and the HPRA.

TARGET GROUPS

Hospital Managers / CEOs	Clinical Directors
Supplies Managers	Purchasing Managers
Risk Managers	Clinical/Biomedical Engineers
All Nursing Staff	Public and Private Nursing Homes
Relevant Wards	Medical Device Distributors

BACKGROUND

Through customer complaints Linet has become aware of improper use of the locking function of the height adjustment mechanism of the bed. When the bed is being used by a patient, such improper use, may cause unintended activation of the bed and in the worst case scenario, injury to a patient or user.

MANUFACTURER CONTACT INFORMATION

Enquiries to the **manufacturer** should be addressed to:

Linet spol s.r.o.	Telephone: +420 312 576 451
Želevčice 5,	E-mail: info@linet.cz
274 01 Slaný	
Czech Republic	

HPRA CONTACT INFORMATION

All **adverse incidents** relating to a medical device should be reported to:

Health Products Regulatory Authority	Telephone: +353-1-6764971
Kevin O'Malley House	Fax: +353-1-6344033
Earlsfort Centre	E-mail: devicesafety@hpra.ie
Earlsfort Terrace	Website: www.hpra.ie
Dublin 2	