



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2023FA0001

Date: 01 March 2023

Urgent Field Safety Notice

Advance Micro™ 14 Ultra Low-Profile PTA Balloon Catheter Ultrathane Cook-Cope Type Locking Loop Multipurpose Drainage Catheter

For Attention of: Chief Executive / Risk Management / Purchasing

Contact details of local representative (name, e-mail, telephone, address etc.)

Cook Medical Europe Ltd.
O'Halloran Road
National Technology Park
Limerick, Ireland
E-mail: European.FieldAction@CookMedical.com
Phone: Please refer to the attached Country Contacts List

For any further information or support concerning the information within this FSN, please contact your local Cook Medical Sales Representative or Cook Medical Europe Ltd.



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2023FA0001

Urgent Field Safety Notice

Advance Micro™ 14 Ultra Low-Profile PTA Balloon Catheter Ultrathane Cook-Cope Type Locking Loop Multipurpose Drainage Catheter

Risk Addressed by FSN

1. Information on Affected Devices	
1.	<p>1. Device Type(s)</p> <p>The Advance Micro 14 Ultra Low-Profile PTA Balloon Catheter is a double-lumen catheter with a balloon near its distal tip. The catheter consists of two independent lumens, which are labelled "DISTAL" and "BALLOON." The distal lumen extends the length of the catheter and is used for placement of wire guides. The balloon lumen is used to expand the balloon. Inscribed on the tip of the manifold are the balloon diameter (mm) and the balloon length (cm). The balloon is manufactured from an extra-thinwall, high-strength, minimally-compliant material.</p> <p>Multipurpose Drainage Catheters are constructed from Ultrathane® or polyethylene and come in a range of French sizes, lengths and sideport quantities.</p>
1.	<p>2. Commercial name(s)</p> <p>Advance Micro™ 14 Ultra Low-Profile PTA Balloon Catheter Ultrathane Cook-Cope Type Locking Loop Multipurpose Drainage Catheter</p>
1.	<p>3. Primary clinical purpose of device(s)</p> <p>The Advance Micro 14 Ultra Low-Profile PTA Balloon Catheter has been designed for percutaneous transluminal angioplasty (PTA) of lesions in peripheral arteries including internal pudendal, iliac, renal, popliteal, femoral, iliofemoral, anterior tibial, posterior tibial, peroneal, pedal, radial, brachial, and ulnar, as well as in obstructive lesions of native or synthetic arteriovenous dialysis fistulae. Not for use in the coronary arteries.</p> <p>Multipurpose drainage catheters are intended for percutaneous drainage in a variety of drainage applications (e.g., nephrostomy, biliary, and abscess), either by direct stick or Seldinger access technique.</p>
1.	<p>4. Device Model/Catalogue/Part Number(s)</p> <p>Reference Part Numbers (RPNs): PTA3-14-150-2.5-8, PTA3-14-150-2.5-3, and ULT16.0-38-25-P-6S-LMCL Order Numbers (GPNs): G26677, G26646, and G07340, respectively</p>
1.	<p>5. Affected serial or lot number range</p> <p>14110999, 14111001, NS13527160, 14012749, and 14007378.</p>



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU


FSN & FSCA Ref: 2023FA0001

2. Reason for Field Safety Corrective Action (FSCA)		
2.	<p>1. Description of the product problem</p> <p>Specific product lots were distributed throughout Europe, Middle East, and Africa (EMEA) after Cook removed the products from the Declaration of Conformity. As such, the products should not have been distributed.</p>	
2.	<p>2. Hazard giving rise to the FSCA</p> <p>There is no health hazard associated with this issue; impacted products conform to manufacturing tolerances and specifications. Products are being removed from the market due to a regulatory/compliance issue.</p>	
3. Type of Action to Mitigate the Risk		
3.	<p>1. Actions To Be Taken by the User</p> <p><input checked="" type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device <input checked="" type="checkbox"/> Other</p> <p>Please complete the enclosed Customer Reply Form. Where product is indicated as being returned, our Customer Services department will contact you to organize the return and issue you with the relevant Returns Authorization number. Please include contact details on the Customer Reply form.</p> <p>Returned Product should be addressed to: Cook Medical EUDC Robert-Koch-Straße, 2 52499 Baesweiler GERMANY</p> <p>Credit will be provided for the returned affected products where applicable.</p>	
3.	<p>2. Is Customer Reply Required?</p> <p>Form is attached specifying deadline for return.</p>	Yes
3.	<p>3. Action Being Taken by the Manufacturer</p> <p><input checked="" type="checkbox"/> Product Removal</p>	



COOK MEDICAL EUROPE LTD.
 O'HALLORAN ROAD
 NATIONAL TECHNOLOGY PARK
 LIMERICK, V94 N8X2, IRELAND
 TEL: +353 61 334440 FAX: +353 61 334441
 WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2023FA0001

4. General Information		
4.	1. FSN Type	Update
4.	2. Further advice or information already expected in follow-up FSN?	No
4.	3. Manufacturer information Refer to page 1 of this FSN for contact details of local representative.	
	a. Company Name	Cook Incorporated
	b. Address	750 Daniels Way Bloomington, IN 47402, United States
4.	4. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.	
4.	5. Name/Signature	
		Larry D. Pool Director, Post Market Cook Incorporated



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2023FA0001

Transmission of this Field Safety Notice

This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred.

Please transfer this notice to other organisations on which this action has an impact.

Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.

Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.