

Atellica® CH Analyzer

Atellica® CI Analyzer

Potential for Negative Bias with Atellica CH Immunoglobulin M₂ (IgM₂) Reagent

Our records indicate that your facility may have received the following product:

Table 1. Atellica CH and Atellica CI Affected Product

Assay	Test Code	Siemens Material Number (SMN)	Unique Device Identification (UDI)	Lot Number
Immunoglobulin M ₂	IgM ₂	11097620	00630414595627	221764 and above

Reason for Correction

The purpose of this communication is to inform you of an issue with the product indicated in Table 1 above and provide instructions on actions that your laboratory must take.

Siemens Healthineers has confirmed the potential for a negative bias with quality control (QC) and patient sample results when using the Atellica CH Immunoglobulin M₂ (IgM₂) reagent. The negative bias was observed after the IgM₂ reagent was stored onboard the analyzer regardless of whether the reagent wells were punctured or unpunctured. Unopened reagents stored refrigerated at 2 - 8 °C are unaffected. See Tables 2 and 3 in the Additional Information section for observed results.

This correction is applicable to all future lots until further notice. Siemens Healthineers is currently investigating the root cause of this issue.

Risk to Health

When this issue occurs, there is a potential for erroneously depressed IgM patient results. This is not expected to lead to a significant effect on assessment of IgM results in the context of the assay intended use.

Actions to be Taken by the Customer

Siemens Healthineers recommends batch testing samples for Atellica CH IgM₂ as follows:

1. Remove and discard any Atellica CH IgM₂ reagent packs onboard the analyzer.
2. Load a **single** fresh Atellica CH IgM₂ reagent pack onto the analyzer.
3. Perform a **Lot calibration** and process Quality Control (QC).
4. Immediately process a batch of patient samples and conclude with a repeat run of QC.

Patient results should not be reported until the QC performed at the end of the batch run has been assessed.

- If the QC results are within the established range, patient results can be reported.
 - If the QC results are not within the established range, do not report patient results and repeat steps 1 – 4 above.
5. Remove and discard the Atellica CH IgM₂ reagent pack at the end of the batch run.
- Siemens Healthineers does not recommend using the ADVIA IgM₂ reagent on the Atellica CH or Atellica CI Analyzers.
 - Please review this letter with your Medical Director to determine the appropriate course of action, including for any previously generated results, if applicable.
 - Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
 - If you have received any complaints of illness or adverse events associated with the product listed in Table 1, immediately contact your local Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

Please retain this letter with your laboratory records and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

Additional Information

Table 2. Atellica CH IgM₂ Quality Control Results Across Four Hours*

Time (hour)	QC Level 1		QC Level 2		QC Level 3	
	mg/dL (g/L)		mg/dL (g/L)		mg/dL (g/L)	
	Rep 1	Rep 2	Rep 1	Rep 2	Rep 1	Rep 2
0 (after calibration)	47.0 (0.47)	45.5 (0.46)	76.3 (0.76)	72.2 (0.72)	90.4 (0.90)	88.9 (0.89)
1	41.6 (0.42)	38.9 (0.39)	72.2 (0.72)	70.1 (0.70)	79.8 (0.80)	84.7 (0.85)
2	38.4 (0.38)	37.1 (0.37)	65.4 (0.65)	64.7 (0.65)	77.5 (0.78)	78.5 (0.79)
3	34.5 (0.35)	33.9 (0.34)	63.3 (0.63)	64.7 (0.65)	75.5 (0.76)	75.4 (0.75)
4	32.2 (0.32)	33.5 (0.34)	63.1 (0.63)	63.1 (0.63)	73.2 (0.73)	75.9 (0.76)

*Quality Control (QC) materials are representative of patient samples. Bio-Rad Multiquel Assayed Control Lot 45960 was used for testing.

Table 3. Atellica CH IgM₂ Patient %Bias Results at 24 Hours**

Range of Patient Serum Sample IgM ₂ Values	%Bias Range	Average %Bias
23 – 50 mg/dL (0.23 - 0.50 g/L)	[-14.9 to -26.5%]	-20.7%
51 – 100 mg/dL (0.51 - 1.00 g/L)	[-6.7 to -15.4%]	-11.0%
101 – 200 mg/dL (1.01 - 2.00 g/L)	[-4.1% to -6.3%]	-5.2%
201 – 330 mg/dL (2.01 - 3.30 g/L)	[-2.9% to -4.7%]	-3.8%

**Initial result was obtained immediately after calibration of a freshly loaded reagent pack. A repeat result was obtained 24 hours after calibration from the same reagent well.

Atellica® is a trademark of Siemens Healthineers.

Potential for Negative Bias with Atellica CH Immunoglobulin M_2 (IgM_2) Reagent

FIELD CORRECTION EFFECTIVENESS CHECK

Potential for Negative Bias with Atellica CH Immunoglobulin M_2 (IgM_2) Reagent

This response form is to confirm receipt of the enclosed Siemens Healthineers Urgent Field Safety Notice ACHC24-01.A.OUS dated December 2023 regarding the Potential for Negative Bias with Atellica CH Immunoglobulin M_2 (IgM_2) Reagent. Please read each question and indicate the appropriate answer.

Return this completed form to Siemens Healthineers as per the instructions provided at the bottom of this page.

1. I have read and understood the UFSN instructions provided in this letter. Yes No
2. Is your laboratory currently running the assay listed in Table 1 on the Atellica Yes No
Analyzer?

Name of person completing questionnaire: _____

Title: _____

Institution: _____ Instrument Serial Number: _____

Street: _____

City: _____ State: _____

Phone: _____ Country: _____

Please send a scanned copy of the completed form via email to cruinnfsngroup@cruinn.ie.

Or fax this completed form to the Customer Care Center at 01 6297401.

If you have any questions, contact your local Siemens Healthineers technical support representative.