

Product

Lot

Expiry

Date of Notification:

Details of problems with this device:

A batch of Nasopure have been identified to contain insert designed for the Danish market instead of English insert. This could lead to potential misuse of the device. As a result, this product is seen to be defective.

Action to be taken by the user:

Danish stock can be identified by a clear sticker that contains Danish text, as per the below image. All stock identified to be 'defective' should be returned to the supplier who will issue replacements.



Advisory Notice

This notice should be distributed to all those who need to be made aware within your organisation and to any organization where the device has been transferred (if appropriate).

Please Maintain awareness on of this notice and resulting action to an appropriate period to ensure effectiveness of this notice.

Issued on behalf of Pharmed Ltd by:

Charlotte Dugdale

Title:

Quality and Compliance/ RP

Contact Details:

Pharmed Group, 3 Clonmore Business Park, Mullingar, Co. Westmeath, Ireland. N91DK59. Quality@pharmed-group.com

Customer – Please return this slip to the address or email above

I confirm that I have received the advisory notice regarding the device below and have taken the required action:

Customer

Name	Signature	Date
Product Name	Lot Number	Expiry